



**Vermont Department of Health
Division of Health Surveillance**
Public Health Laboratory
195 Colchester Avenue - P.O. Box 1125
Burlington, VT 05402-1125

[phone] 802-863-7335
[fax] 802-863-7632
[or] 800-660-9997

Agency of Human Services

RECEIVED

AUG 18 2011

COUNTY: Addison

Page 1 of 1

Addison State's Attorney

August 15, 2011

Dear colleague:

Enclosed is a packet of information that includes all technical information regarding each of the active DataMaster DMT instruments in your county. This documentation would include maintenance and repair records, certification records, and installation records. This would include documents created or received during July 2011.

If you have questions or suggestions regarding this process, please don't hesitate to contact me at (802) 863-7570.

Thank you.

Sincerely,

Mary Celotti
Laboratory Director

Enclosures:

Datamaster DMT S.N. 100137 (Middlebury Police Dept)

No updates on technical information during July 2011.

Datamaster DMT S.N. 100135 (New Haven VSP)

Routine Performance Check Report, 07/05/11

Datamaster DMT S.N. 100141 (Vergennes Police Dept)

No updates on technical information during July 2011


Datamaster DMT S.N. 100134 (Bristol Police Dept)

Routine Performance Check Report with 4 attachments, 07/01/11



ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 100135
Location: NEW HAVEN VSP
Calibration Date: 04/27/2010
Certification Date: 04/27/2010
Installation Date: 04/28/2010
RPC Date: 07/05/2011
Supervisor Name: STEPHANIE J TAYLOR



Diagnostic Results

VERSIONS
DMT 1.04
PIC 2.05
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES

Sample Chamber = 49.09°C
Breath Tube = 48.07°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.83 V
Cooler Voltage = 1.71 V
Bias Voltage = 80.8 V
Chopper Freq = 552.9 Hz

PUMP INFO

Flow Rate = 5.991 L/M

DETECTOR INFO

PUMP ON OFF
MAX(V) 0.2083 0.2110
MIN(V) 0.2065 0.2093

FILTER INFO

Filter 1 0.210 Zero = true
Filter 2 0.486 Zero = true
Filter 3 0.778 Zero = true

CALIBRATION CHECK

Xq = 0.0872 0.30%

Routine Performance Check Passed

Accuracy and Precision Check


Concentration = 0.103 g/210L
Lot # = 11-19-100
Average = 0.105 g/210L
Std Dev = 0.000

RFI Interference Test

RFI detected /

Sample Acceptance Test

Passed /



Performed by Stephanie Taylor

Date 07/05/2011

Reviewed by Steven Harnovs

Date 7/7/11

DMT Serial Number #100135

Page 1 of 1

07/05/2011 18:08

Reviewed by AKK 7-12-11

DMT -- 100135-000058